
A Laser Based Semi-Autonomous Wheelchair Using Bayesian Theory

By

Tuyen Hoang Trieu

Submitted to Faculty of Engineering and Information Technology

in partial fulfilment of Doctor of Philosophy

at the University of Technology, Sydney.


Sydney, September 2009

Certificate of Authorship/Originality

I, Tuyen Hoang Trieu, certify that the work in this thesis has not previously been submitted for a degree, nor has it been submitted as part of the requirements of a degree, except as fully acknowledged within the text.

I also certify that this thesis has been written by me. Any help that I have received in my research work and in the preparation of the thesis itself has been acknowledged. In addition, I certify that all information sources and literature used are indicated in the thesis.

Signature of Candidate

A handwritten signature in black ink, appearing to read 'Hoang', followed by a long, sweeping horizontal stroke.

Sydney, September 2009.

Acknowledgements

First of all, I would like to thank my supervisor, Professor Hung Tan Nguyen, for giving me the great opportunity to work on the smart wheelchair project under his expert guidance. He is not only my supervisor, but has also been like a father to me, who has supported me through this critical period of my study. I am very grateful for his support and guidance.

I would also like to thank my co-supervisor, Dr Keith Willey, for his valuable help throughout this project, from the first semester until the very end of my journey.

Special thanks to my wife for her great encouragement and support. She is my best friend, my soul mate and my family.

Many thanks to Samantha Yee for her help in editing this thesis.

Finally, this thesis is dedicated to my beloved mother, my father and my younger brother.

Tuyen Hoang Trieu

Sydney, September 2009

Contents

List of Figures VII

List of Tables XI

Chapter 1 Introduction 6

 1.1. Problem Statement 7

 1.2. Objectives of the thesis 9

 1.3. Thesis contributions 10

 1.4. The structure of the thesis 13

 1.5. Publications related to the thesis 15

Chapter 2 Literature Review 17

 2.1. Introduction 17

 2.2. Autonomous Systems 19

 2.2.1. Sensor Introduction 19

 2.2.1.1. Common Range Sensors and Limitations 19

 2.2.1.2. Laser Range Finder Sensors and Advances 22

2.2.2.	Obstacle Avoidance Techniques	24
2.2.2.1.	Obstacle Avoidance Techniques Based on Prior Environmental Information	25
2.2.2.2.	Obstacle Avoidance Techniques Based on Online Environmental Information	26
2.3.	Smart Wheelchairs and Shared Control Strategies	29
2.3.1.	User Interfaces	29
2.3.2.	Smart Wheelchair Review	34
2.3.2.1.	Smart Wheelchair Based on User's Control Capability	34
2.3.2.2.	Smart Wheelchairs Requiring Prior Knowledge of Environments	37
2.3.2.3.	Smart Wheelchairs Based on Pre-designed Autonomous Obstacle Avoidance Tasks	41
2.4.	Discussion	50
Chapter 3	Real-time Mapping and Free-space Estimation for a Laser Based Wheelchair	53
3.1.	Introduction	53
3.2.	SAM – Smart Wheelchair System	55
3.3.	Laser Range Finder System	57
3.4.	Methodology	61
3.4.1.	Real-time Laser Map	61
3.4.1.1.	Position Estimation – The Wheelchair Model	62
3.4.1.2.	Construction of a Real-time Laser Map	65
3.4.1.3.	UMBmark - Calibration Procedure	70

3.4.2.	Available Accessible Space	73
3.4.2.1.	A Survey of Contemporary Methods	73
3.4.2.2.	Advanced Method of Determining Accessible Free-space	75
3.5.	Experimental Results	78
3.5.1.	Real-time 2D Mapping Result	78
3.5.2.	The Calibration and Results	80
3.5.3.	Accessible Free-space Results	86
3.6.	Discussion	88

Chapter 4 Advanced Obstacle Avoidance Based on Bayesian Neural

	Networks for a Laser Based Wheelchair System	90
4.1.	Introduction	90
4.2.	Bayesian Neural Networks	92
4.2.1.	Standard Neural Networks	92
4.2.2.	Why Bayesian Neural Network?	95
4.2.3.	Bayesian Framework for Feed-Forward Neural Networks	97
4.3.	Advanced Obstacle Avoidance Based on Bayesian Neural Networks – Methodology	100
4.3.1.	Training Algorithm for Obstacle Avoidance Bayesian Neural Networks	100
4.3.2.	Hessian Matrix for Evidence Estimation	103
4.3.3.	Bayesian Framework for Obstacle Avoidance	105
4.3.4.	Obstacle Avoidance Neural Networks	108
4.3.5.	Data Acquisition Method	112
4.4.	Training and Experimental Results	113

4.4.1.	Data Acquisition Results	113
4.4.1.1.	General Obstacle Avoidance Data Acquisition Task	113
4.4.1.2.	Corridor and Wall Following Data Acquisition Task	114
4.4.1.3.	Door Passing Data Acquisition Task	115
4.4.2.	Training Conditions	116
4.4.3.	Network Training and Experimental Results	119
4.4.3.1.	General Obstacle Avoidance Task (GOA)	119
4.4.3.2.	Corridor and Wall Following Task (C-WF)	123
4.4.3.3.	Door Passing Task (DP)	126
4.5.	Discussion	129
Chapter 5	Shared Control Strategy Based on Bayesian Theory for an	
	Intelligent Laser Based Wheelchair System	131
5.1.	Introduction	131
5.2.	Control Structure for the SAM Assistive Wheelchair	133
5.3.	Shared Control Strategy Based on the Bayesian Recursive	
	Technique - Methodology	135
5.3.1.	Recursive Bayesian Technique	138
5.3.2.	Bayesian Framework for the Adaptive Shared Control Strategy	140
5.3.3.	Environment Model	145
5.4.	The Experimental Results of the Shared Control Strategy	146
5.4.1.	Training and Testing Result of Environmental Model	147
5.4.2.	Shared Control Strategy – Experimental Results	153
5.5.	System Evaluations	163
5.5.1.	Task 1 – Navigating in a Narrow Space Experiment	164

5.5.2.	Task 2 – Corridor Following and Door Passing Experiment	168
5.5.3.	Task 3 – Navigate Through a Tight Spot Experiment	172
5.6.	Discussion and Conclusion	176
Chapter 6	Conclusions and Future Work	178
6.1.	Discussions and Conclusions	178
6.2.	Future Work	181

Bibliography

Appendix A	Bayesian Framework for Neural Networks	191
A.1.	Determining the Optimal Network's Weights	191
A.1.1.	Gaussian Approximation of Posterior.....	195
A.1.2.	Gaussian Approximation of Likelihood	196
A.1.3.	Gaussian Approximation of Prior	196
A.1.4.	Estimating the Evidence of Hyper-Parameters	197
A.1.5.	Estimating the Most Probable Hyper-Parameters	200
A.2.	Determining the Optimal Network's structure	201
A.2.1.	Occam's razor – Bayesian penalisation factor	201
A.2.2.	Model comparison – selecting the most optimal network' structure	205
Appendix B	The Wheelchair's Performance Evaluation Result	208
B.1.	The 1st Experimenter's testing results.	209
B.2.	The 2nd experimenter's testing results.	212
B.3.	The 3rd experimenter's testing results.	215

B.4.	The 4th experimenter’s testing results.	218
B.5.	The 5th experimenter’s testing results.	221
B.6.	The 6th experimenter’s testing results.	224
B.7.	The 7th experimenter’s testing results.	227
B.8.	The 8th experimenter’s testing results.	230
Appendix C Software Code		233
C.1.	Bayesian Training – Matlab Code	233
C.1.1.	Basic Training Code	233
C.1.2.	Bayesian training with Levenberg-Marquardt algorithm	234
C.1.3.	AlphXI Function	243
C.2.	Control Software – C Code	244
C.2.1.	WC function library.h	245
C.2.2.	WC function library.c	255

List of Figures

Figure 2.1: An overall structure of a smart wheelchair control system.	12
Figure 2.2: Typical range sensors: (A) - Sonar sensor; (B) – Infrared sensor.	14
Figure 2.3: Bumblebee®2 camera system	15
Figure 2.4: Examples of laser range finder systems	17
Figure 2.5: Head-movement interface	24
Figure 2.6: EEG control interface	25
Figure 2.7:Hephaestus smart wheelchair	28
Figure 2.8: CWA autonomous wheelchair	32
Figure 2.9: Robchair prototype	33
Figure 2.10: SENA wheelchair	34
Figure 2.11: SMARTCHAIR prototype	37
Figure 2.12: Osaka University’s wheelchair	39
Figure 2.13: The MAid wheelchair	40
Figure 2.14: The NavChair prototype	42
Figure 3.1: SAM intelligent wheelchair	50
Figure 3.2: The SAM wheelchair’s hardware overview	50
Figure 3.3: URG 04-LX laser sensor	52
Figure 3.4: Data acquisition software for URG 04-LX laser sensor.	53
Figure 3.5: 3D laser hardware system.	54
Figure 3.6: The software for data acquisition of 3D laser sensor	54

Figure 3.7: Illustrations of wheelchair's movement	57
Figure 3.8: Relative co-ordinates for the SAM wheelchair	60
Figure 3.9: Updating positions of surrounding obstacles	60
Figure 3.10: Five steps of building a real-time local map	62
Figure 3.11: The Bi-directional Square Path experiment, UMBmark method (Borenstein & Feng 1996)	65
Figure 3.12: The accessible free-space estimating procedure.	70
Figure 3.13: The algorithm of determining the accessible free-space from the real-time 2D laser map.	71
Figure 3.14: The results of the real-time 2D local map method.	73
Figure 3.15: The wheelchair calibration software	75
Figure 3.16: The results of UMBmark experiments of the uncalibrated wheelchair..	76
Figure 3.17: The results of UMBmark experiments after UMBmark calibration. .	78
Figure 3.18: The trajectories of the UMBmark experiments.	79
Figure 3.19: The results of the 2D mapping method for	79
Figure 3.20: The results of determining accessible free-space for SAM	81
Figure 4.1: Structure of a two perceptron layered feed-forward neural network.....	87
Figure 4.2: Ockham's razor mechanism of Bayesian framework (Bishop 1995)	90
Figure 4.3: The algorithm of estimating the Hessian matrix.	99
Figure 4.4: The Bayesian framework for the obstacle avoidance task	101
Figure 4.5: The structure of Bayesian neural networks for obstacle avoidance.	104
Figure 4.6: The comparison of training performance between Levenberg-Marquardt and backpropagation algorithm.	111

Figure 4.7: The training and testing results for the general obstacle avoidance task 114

Figure 4.8: Experimental result of Bayesian neural network in performing
the general obstacle avoidance task. 115

Figure 4.9: The training and testing results for the corridor and
wall following task 118

Figure 4.10: Experimental results of Bayesian neural network in performing
the corridor and wall following task. 119

Figure 4.11: The training and testing results for the door passing task 121

Figure 4.12: Experimental results of Bayesian neural network in
performing the door passing task. 122

Figure 5.1: Overall control diagram of the assistive wheelchair system. 128

Figure 5.2: The Markov chain for the wheelchair application 135

Figure 5.4: The structure of the adaptive shared control strategy 139

Figure 5.5: Environmental classifier. 140

Figure 5.6: Real-time control software for the smart wheelchair. 141

Figure 5.7: Training result for environment classification task. 143

Figure 5.8: Experimental results of testing environment model. 146

Figure 5.9: The performance of the adaptive shared control strategy based on the
recursive Bayesian technique. 150

Figure 5.10: The performance of the adaptive shared control strategy based on the
recursive Bayesian technique 151

Figure 5.11: The evidences of the three autonomous tasks. 152

Figure 5.12: The performances of the two shared control strategies
in the first experiment. 154

Figure 5.13: The performances of the two shared control strategies are compared
in the second experiment. 155

Figure 5.14: The performances of the two shared control methods in the second
experiment (continued). 156

Figure 5.15: The environment for the ‘navigating to a narrow area’ experiment 158

Figure 5.16: The result of experimenter 3 in Task 1 159

Figure 5.17: Time for finishing Task 1 for 8 experimenters in manual
and semi-autonomous modes. 161

Figure 5.18: Average time for manual and semi-autonomous modes 161

Figure 5.19: The environment for the ‘corridor following and
door passing’ experiment (Task 2). 162

Figure 5.20: The result of experimenter 1 in Task 2 163

Figure 5.21: Time of finishing Task 2 for 8 experimenters in manual and
semi-autonomous modes. 165

Figure 5.22: Average time for manual and semi-autonomous modes (Task 2) 165

Figure 5.23: The environment for the ‘tight spot’ experiment (Task 3). 166

Figure 5.24: The result of experimenter 8 in Task 3 167

Figure 5.25: Time for finishing Task 3 for 8 experimenters in manual and
semi-autonomous modes. 169

Figure 5.26: Average time in manual and semi-autonomous modes (Task 3) 170

List of Tables

TABLE 3.1:	HOKUYO LASER RANGE FINDER SYSTEM SPECIFICATIONS	52
TABLE 4.1:	DATA ACQUISITION RESULTS OF GENERAL OBSTACLE AVOIDANCE TASK	108
TABLE 4.2:	DATA ACQUISITION RESULTS OF CORRIDOR AND WALL FOLLOWING TASK	109
TABLE 4.3:	DATA ACQUISITION RESULTS OF DOOR PASSING TASK	110
TABLE 5.1:	EXTRACTED LASER IMAGE SAMPLES	143
TABLE 5.2:	RESULTS FOR THE TEST SET USING THE DATA OF DYNAMIC WINDOWS FROM THE ENVIRONMENTS 1 & 3	145
TABLE 5.3:	RESULT FOR THE VERIFICATION SET USING THE LASER IMAGES FROM THE ENVIRONMENT 2	145
TABLE 5.4:	TAKE INTO ACCOUNT ENVIRONMENT MODEL	148
TABLE 5.5:	PROBABILITIES OF CHANGING THE TASK TO GOA	148
TABLE 5.6:	PROBABILITIES OF CHANGING THE TASK TO CWF	149
TABLE 5.7:	PROBABILITIES OF CHANGING THE TASK TO DP	149
TABLE 5.8:	INFORMATION FROM EXPERIMENTS	158
TABLE 5.9:	SUMMARY OF TIME TAKEN FOR EXPERIMENT 1	160
TABLE 5.10:	SUMMARY OF EXPERIMENTAL RESULTS OF TASK 2	164
TABLE 5.11:	SUMMARY OF EXPERIMENTAL RESULTS OF TASK 3	169

Abstract

Reports show that the number of elderly people and people with disabilities in society are significantly increasing. These people have a wide range and variety of functional impairments. The aim of rehabilitation technology is to improve the quality of life of people with disabilities. In particular, smart wheelchairs have been developed to accommodate people with mobility impairments. The provision of independent mobility can produce substantial benefits, such as the development of physical, cognitive, communication and social skills for both children and adults in their daily lives.

In this thesis, we present a method of constructing a 360° real-time environmental map for the smart wheelchair. It combines the information from a laser range finder and encoders mounted in the driving wheels. As the wheelchair moves, the obstacles in front of the wheelchair that are detected by the laser sensor are updated to this map, after a modification based on the encoder data. This mapping method is 13-fold more accurate than the common use of encoders.

Also, a method of determining accessible free-space for the obstacle avoidance task is implemented. This accommodates the actual dimensions of the wheelchair and then determines the collision-free area available for the wheelchair. The data of accessible free-space can be used to simplify the obstacle avoidance controller and improve its real-time capabilities. It eliminates errors from the common approximation of the wheelchair boundary as a circle, which may lead to the assumption that no accessible space is available to move safely without incurring a collision.

We introduce an advanced obstacle avoidance technique that utilises separate Bayesian neural networks for specified tasks. The obstacle avoidance task is divided into three sub-tasks: passing through a door, corridor and wall following, and general obstacle avoidance. This enables the network to respond to the particular features of each task, thereby improving performance. Specific data acquisitions are performed to collect the patterns used to design the neural network for each task. Bayesian framework is then applied to determine the optimal network structures. The training patterns are subsequently used in conjunction with the Bayesian training process to improve the generalisation and performances of each network. Our method was able to successfully accomplish difficult navigation tasks smoothly following a near optimum trajectory.

Furthermore, we developed an adaptive shared control method for an intelligent wheelchair based on the Bayesian recursive technique to assist disabled users when performing obstacle avoidance tasks. Three autonomous tasks have been developed for different types of environments to improve the performance of the overall system. The system combines local environmental information gathered using a laser range finder sensor with the user's intentions to select the most suitable autonomous task in different situations. The evidences of these tasks are estimated by the Bayesian recursive technique during movements of the wheelchair. The most appropriate task that is selected automatically by the wheelchair controller is the one that has the highest value of evidence.

Finally, a method of classifying the environmental model is introduced for this shared control strategy. The features of the environment such as doorways, corridors and walls, and general obstacles have to be recognised. This method is based on the Bayesian neural network to recognise the environmental features from the laser images that were acquired from the onboard sensors. This environmental feature information is one of the main inputs of the adaptive shared control strategy to effectively improve the accuracy of autonomous task selection.

Various experiments are conducted to evaluate the performance of our smart wheelchair system. Eight able-bodied people are recruited, including four males and four females whose ages range from 27 years to 60 years. They have had no experience of driving a

wheelchair before. Three experiments are arranged with increasing difficulties. These users are asked to drive the wheelchair in both manual mode via a conventional joystick and in semi-autonomous mode with the shared control strategy and autonomous obstacle avoidance controller.

The evaluation results show the advantages of our wheelchair's control system compared to the manual control method. Our intelligent wheelchair, which was equipped with a reliable obstacle avoidance method and an effective shared control strategy, was able to successfully accomplish difficult navigation tasks not only by following a near optimum trajectory but also by generating smooth movements (maintaining stable velocities) in different types of environments. It is able to successfully support people with various types and levels of impairment.